






Chlamydia and GC Collection Devices

	<p>CT/GC BD Qx Endocervical Female Swab Female Endocervical Specimen Collection Kit for the BD Probe Tec™ <i>Chlamydia trachomatis/Neisseria gonorrhoeae</i> (CT/GC) Q^x Amplified DNA Assays. Includes: Cleaning swab Discard white cleaning swab after use</p> <ul style="list-style-type: none">• White shaft• Sterile fiber (Rayon) tip <p>Sterile collection swab Use pink swab for collection</p> <ul style="list-style-type: none">• Scored pink plastic shaft• Sterile polyurethane foam tip <p>Swab Diluent for the BD Probe Tec™ CT/GC Q^x Amplified DNA Assays</p> <ul style="list-style-type: none">• Black pierceable cap <p>Storage and handling requirements: 2° to 33°C</p>
	<p>CT/GC BD Qx Urethral Swab Male Urethral Specimen Collection Kit for the BD Probe Tec™ <i>Chlamydia trachomatis/Neisseria gonorrhoeae</i> (CT/GC) Q^x Amplified DNA Assays. Includes: Sterile collection swab Use blue swab for collection</p> <ul style="list-style-type: none">• Scored blue plastic shaft• Stainless steel wire• Sterile rayon tip <p>Swab Diluent for the BD Probe Tec™ CT/GC Q^x Amplified DNA Assays</p> <ul style="list-style-type: none">• Black pierceable cap <p>Storage and handling requirements: 2° to 33°C</p>
	<p>CT/GC BD Qx Vaginal Specimen Transport Vaginal Swab <i>Chlamydia trachomatis/Neisseria gonorrhoeae</i> (CT/GC) Q^x Amplified DNA Assays</p> <p>Swab: Dry Transport</p> <ul style="list-style-type: none">• Transport sheath• Sterile collection swab• Purple plastic shaft• Sterile polyester tip• <p>Storage and handling requirements: 2° to 30°C</p>

Chlamydia and GC Collection Devices



CT/GC BD Qx Urine Preservative Transport (UPT) tubes

Urine Preservative tubes for the BD Probe Tec™ *Chlamydia trachomatis/Neisseria gonorrhoeae* (CT/GC) Q^x Amplified DNA Assays. Includes:

- UPT tubes
- Transfer pipettes

Storage and handling requirements: 2° to 33°C